

ORGANIC - SOIL				
PARAMETER	CONT. TYPE	HOLD TIME	BOTTLE SIZE	PRESERVATIVE
BTEX/PIRI C6 - C10/CCME F1/VOCs	CV	40 days	2 x 40 mL	MeOH
PIRI C10 - C32/CCME F2-F4	CG/AG	14 days	120 mL	None
PAH/OCP	AG	14 days	120 mL	None
PCB	CG/AG	None	120 mL	None

INORGANIC - SOIL				
PARAMETER	CONT. TYPE	HOLD TIME	BOTTLE SIZE	PRESERVATIVE
Ammonia	CG	28 days	120 mL	None
Anions	CG	28 days	120 mL	None
Cations	CG	6 months	120 mL	None
Conductivity	CG	28 days	120 mL	None
Cyanide	CG	14 days	120 mL	None
Hexavalent Chromium	CG	28 days	120 mL	None
MeHg	CG	6 months	120 mL	None
Mercury	CG	3 months	120 mL	None
Metals	CG	3 months	120 mL	None
Nitrate & Nitrite	CG	7 days	120 mL	None
pH	CG	3 months	120 mL	None
Phenols	CG	28 days	120 mL	None
Total Organic Carbon (TOC)	CG	14 days	120 mL	None

BIOLOGICAL - WATER				
PARAMETER	CONT. TYPE	HOLD TIME	BOTTLE SIZE	PRES.
Biological Oxygen Demand (BOD)	PB	48 hours	500 mL	None
Fecal Coliforms MF or MPN	PB	24 hours	250 mL	Na ₂ S ₂ O ₃
<i>Pseudomonas aeruginosa</i> presumptive	PB	24 hours	250 mL	Na ₂ S ₂ O ₃
Heterotrophic Plate Count (HPC)	PB	24 hours	250 mL	Na ₂ S ₂ O ₃
Total Coliforms & E.coli MPN or MF	PB	24 hours	250 mL	Na ₂ S ₂ O ₃
Total Coliforms & E.coli (P/A)	PB	24 hours	250 mL	Na ₂ S ₂ O ₃

BIOLOGICAL - SOIL				
PARAMETER	CONT. TYPE	HOLD TIME	BOTTLE SIZE	PRES.
Fecal Coliforms	250 mL PET	48 hours	100 g	None
Salmonella	250 mL PET	48 hours	100 g	None
E. Coli	250 mL PET	48 hours	100 g	None
Total Coliforms	250 mL PET	48 hours	100 g	None

Please Note: Biological samples will not be accepted after 2 pm on Fridays

Directions: Locate the desired Test. Identify the sample type, container, preservative and holding times. Keep Samples cool and ship to the lab as soon as possible. Pre-cleaned containers can be supplied by the lab.

Please contact your laboratory for more detailed sampling instructions or to notify of incoming RUSH samples.

CONTAINER CODES

AG/AV - Amber Glass/Vial NH - No Headspace TL - Teflon Lined PB - Plastic Bottle CG/CV - Clear Glass/Vial

ORGANIC - WATER				
PARAMETER	CONT. TYPE	HOLD TIME	MIN. AMOUNT	PRESERVATIVE
BTEX/PIRI C6 - C10/CCME F1/VOCs	CV/AV, NH	14 days	3 x 40 mL	NaHSO ₄
CCME F2-F4	CG/AG	14 days	2 x 100 mL	HCl
Dissolved Gases	CV/AV, NH	14 days	3 X 40 mL	HCl
PAH	CG/AG	14 days	2 x 100 mL	NaHSO ₄
Low Level PAH	CG/AG	14 days	2 x 1L	NaHSO ₄
PIRI C10 - C32	CG/AG	14 days	2 x 250 mL	HCl
PCB	CG/AG	None	2 X 250 mL	None
TOG/MOG/F4G/AV	CG	28 days	2 x 1L	H ₂ SO ₄

WATER - CHLORINATED SOURCE				
HAA	AV, NH	14 days	3 x 40 mL	NH ₄ Cl
THM	AV, NH	14 days	3 x 40 mL	Na ₂ S ₂ O ₃

INORGANIC - WATER				
PARAMETER	CONT. TYPE	HOLD TIME	MIN. AMOUNT	PRESERVATIVE
Ammonia	PB	28 days	120 mL	H ₂ SO ₄
Anions	PB	28 days	120 mL	None
Cations	PB	28 days	120 mL	HNO ₃
Chemical Oxygen Demand (COD)	PB, NH	28 days	120 mL	H ₂ SO ₄
Chlorine, Residual (Total or Free), Chloramines	AV	ASAP	2 x 40 mL	None
DO, ORP, pH	PB	ASAP	120 mL	None
Colour	PB	7 days	120 mL	None
Cyanide	PB	7 days	60 mL	NaOH
Dissolved Organic Carbon (DOC)	PB	28 days	120 mL	H ₂ SO ₄ after filtering
Ferrous Iron	CV/AV, NH	7 days	2 x 40 mL	HCl after filtering
Hexavalent Chromium	PB	28 days	60 mL	(NH ₄) ₂ SO ₄ + NaOH
MeHg (Dissolved)	FLPE	48 hours (unless filtered and frozen)	250 mL	Pres and filtration completed at the lab
MeHg (Total)	FLPE	48 hours (unless frozen)	250 mL	Pres completed at the lab
Mercury (Dissolved)	AV	28 days	2 x 40 mL	HCl after filtering
Mercury (Total)	AV	28 days	2 x 40 mL	HCl
Metals (Dissolved)	PB	28 days	120 mL	HNO ₃ after filtering
Metals (Total)	PB	28 days	120 mL	HNO ₃
Nitrate & Nitrite	PB	5 days	120 mL	None
Phenols	AG	28 days	120 mL	H ₂ SO ₄
Phosphorus (Total)	PB	28 days	120 mL	H ₂ SO ₄
Standard Water Analysis	PB	7 days **	250 mL, 120 mL, 120 mL	None, H ₂ SO ₄ , HNO ₃
Sulfide	PB	28 days	60 mL	NaOH + ZnAc
Total Dissolved Solids (TDS)	PB	7 days	500 mL	None
Total Inorganic Carbon (TIC)	PB, NH	28 days	120 mL	None
Total Kjeldahl Nitrogen (TKN)	PB	28 days	120 mL	H ₂ SO ₄
Total Organic Carbon (TOC)	PB	28 days	120 mL	H ₂ SO ₄
Total Suspended Solids (TSS)	PB	7 days	500 mL	None
Turbidity	PB	5 days	120 mL	None
Ultra Low Level Hg (Total/Dissolved)	FLPE	28 days	250 mL	None
Ultra Low Level TSS	PB	7 days	1L	None

** Excluding Turbidity, Nitrate + Nitrite, which have 5 day holding times